



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

ELLAB, INC.
 4030 Skyron Drive, Unit D
 Doylestown, PA 18902
 Anthony Pileggi Phone: 215 345 4900

CALIBRATION

Valid To: January 31, 2026

Certificate Number: 4828.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations^{1,7}:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
DC Voltage ³ – Measure Auto, 10 MΩ, 1 MΩ Auto, 10 MΩ, 1 MΩ Auto, 10 MΩ, 1 MΩ Auto, 10 MΩ 1 MΩ Auto, 10 MΩ 1 MΩ	Up to 200 mV 200 mV to 2 V (2 to 20) V (20 to 200) V (20 to 200) V 200 V to 1.0 kV 200 V to 1.0 kV	8.9 μV/V + 0.40 μV 9.8 μV/V + 0.60 μV 2.8 μV/V + 1.0 μV 4.1 μV/V + 61 μV 9.0 μV/V + 61 μV 4.3 μV/V + 530 μV 9.1 μV/V + 26 mV	Fluke 8588A
DC Voltage ³ – Generate	(0 to 220) mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	6 μV/V + 0.40 μV 4 μV/V + 0.7 μV 3 μV/V + 2.5 μV 2.8 μV/V + 4 μV 4 μV/V + 40 μV 5.2 μV/V + 400 μV	Fluke 5730A
DC Current ³ – Measure	Up to 20 μA (20 to 200) μA 200 μA to 2 mA (2 to 20) mA (20 to 200) mA 200 mA to 2 A (2 to 20) A (20 to 30) A	24 μA/A + 0.4 nA 9 μA/A + 0.4 nA 9 μA/A + 10 nA 24 μA/A + 0.1 μA 33 μA/A + 2 μA 100 μA/A + 0.2 mA 180 μA/A + 0.8 mA 500 μA/A + 4.4 mA	Fluke 8588A

Parameter/Equipment	Range	CMC ^{2, 5} (±)	Comments
DC Current ³ – Generate	(0 to 220) µA (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A (2.2 to 3.1) A (3.1 to 12) A (12 to 30.2) A	31 µA/A + 6 nA 27 µA/A + 7 nA 27 µA/A + 40 nA 35 µA/A + 0.7 µA 67 µA/A + 12 µA 230 µA/A + 150 µA 230 µA/A + 250 µA 780 µA/A + 500 µA	Fluke 5730A Fluke 5560A
Clamp On Only Non Toroidal	(30.2 to 60) A (60 to 1510) A	1.5 % + 0.5 A 0.55 % + 0.5 A	Fluke 5500A coil & Fluke 5560A
DC Resistance ³ – Measure	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω 200 Ω to 2 kΩ (2 to 20) kΩ (20 to 200) kΩ 200 kΩ to 2 MΩ (2 to 20) MΩ (20 to 200) MΩ 200 MΩ to 2 GΩ	15 µΩ/Ω + 8.1 µΩ 12 µΩ/Ω + 28 µΩ 9.0 µΩ/Ω + 100 µΩ 7.0 µΩ/Ω + 1.0 mΩ 7.0 µΩ/Ω + 10 mΩ 7.0 µΩ/Ω + 100 mΩ 11 µΩ/Ω + 20 Ω 23 µΩ/Ω + 200 Ω 43 µΩ/Ω + 20 kΩ 580 µΩ/Ω + 2.0 MΩ	Fluke 8588A
Electrical Calibration of RTDs ³ – Simulation & Measure			
Pt 385, 100 Ω	(-200 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C	0.04 °C 0.05 °C 0.07 °C 0.08 °C 0.09 °C 0.18 °C	Fluke 5560A
Pt 3926, 100 Ω	(-200 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C	0.04 °C 0.05 °C 0.07 °C 0.08 °C 0.09 °C	
Pt 3916, 100 Ω	(-200 to -190) °C (-190 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.19 °C 0.03 °C 0.04 °C 0.05 °C 0.05 °C 0.06 °C 0.07 °C 0.08 °C 0.18 °C	

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
DC Resistance ³ – Generate	0 Ω (0 to 1) Ω (1 to 1.9) Ω 1.9 Ω to 10 Ω (10 to 19) Ω (19 to 100) Ω (100 to 190) Ω 190 Ω to 1 kΩ (1 to 1.9) kΩ (1.9 to 10) kΩ (10 to 19) kΩ (19 to 100) kΩ (100 to 190) kΩ 190 kΩ to 1.0MΩ (1.0 to 1.9) MΩ (1.9 to 10) MΩ (10 to 19) MΩ (19 to 100) MΩ	31 μΩ 74 μΩ 74 μΩ 20 μΩ 18 μΩ 8 μΩ 8 μΩ 5 μΩ 5 μΩ 5 μΩ 5 μΩ 9 μΩ 7 μΩ 12 μΩ 15 μΩ 40 μΩ 68 μΩ 180 μΩ	Fluke 5730A
Electrical Calibration of Thermocouples ³ –Simulation & Measure			
Type E	(-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1000) °C	0.31 °C 0.11 °C 0.09 °C 0.12 °C 0.16 °C	Fluke 5560A
Type J	(-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C	0.19 °C 0.10 °C 0.09 °C 0.11 °C 0.16 °C	
Type K	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1000) °C (1000 to 1372) °C	0.22 °C 0.13 °C 0.12 °C 0.16 °C 0.27 °C	
Type N	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1300) °C	0.26 °C 0.12 °C 0.09 °C 0.09 °C 0.16 °C	
Type R	(0 to 250) °C (250 to 400) °C (400 to 1000) °C (1000 to 1767) °C	0.40 °C 0.23 °C 0.21 °C 0.26 °C	

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical Simulation of Thermocouples ³ – (cont)			
Type S	(0 to 250) °C (250 to 1000) °C (1000 to 1400) °C (1400 to 1767) °C	0.33 °C 0.24 °C 0.25 °C 0.32 °C	Fluke 5560A
Type T	(-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C	0.47 °C 0.16 °C 0.10 °C 0.09 °C	

Parameter/Range	Frequency	CMC ^{2,5} (±)	Comments
AC Voltage ³ – Measure			
Up to 12.12 mV, Auto, 10 MΩ, 1 MΩ	(1 to 2) kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	2200 μV/V + 1.3 μV 1000 μV/V + 1.3 μV 350 μV/V + 1.3 μV 0.3 % + 1.3 μV 1.0 % + 4.8 μV 2.0 % + 4.8 μV	Fluke 8588A
(12.12 to 121.2) mV, Auto, 10 MΩ, 1 MΩ	(1 to 2) kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	3700 μV/V + 0.61 μV 3700 μV/V + 0.61 μV 220 μV/V + 1.2 μV 520 μV/V + 6.1 μV 0.2 % + 36 μV 1.0 % + 120 μV 1.5 % + 610 μV 4.0 % + 1.2 mV 8.0 % + 1.2 mV 15 % + 1.2 mV	
(0.121 to 1.212) V, Auto, 10 MΩ, 1 MΩ	(1 to 2) kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	1900 μV/V + 6.1 μV 1900 μV/V + 6.1 μV 300 μV/V + 12 μV 600 μV/V + 61 μV 0.2 % + 370 μV 1.0 % + 1.3 mV 2.0 % + 6.0 mV 4.0 % + 12 mV 8.0 % + 12 mV 15 % + 12 mV	

Parameter/Range	Frequency	CMC ^{2,5} (±)	Comments
AC Voltage ³ – Measure (cont)			
(1.212 to 12.12) V, Auto, 10 MΩ, 1 MΩ	(1 to 2) kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	1900 μV/V + 60 μV 1900 μV/V + 60 μV 300 μV/V + 120 μV 600 μV/V + 600 μV 0.2 % + 3.6 mV 1.0 % + 12 mV 2.0 % + 60 mV 4.0 % + 120 mV 8.0 % + 120 mV 15 % + 120 mV	Fluke 8588A
(12.12 to 121.2) V, 10 MΩ	1 Hz to 1 kHz (1 to 2) kHz (2 to 10) kHz	1400 μV/V + 0.60 mV 1000 μV/V + 0.60 mV 2 % + 1.2 mV	
(12.12 to 121.2) V, Auto, 1 MΩ	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	160 μV/V + 0.6 mV 170 μV/V + 0.6 mV 230 μV/V + 1.2 mV 540 μV/V + 6.1 mV 170 μV/V + 61 mV 170 μV/V + 600 mV	
(121.2 to 1050) V, 10 MΩ	1 Hz to 1 kHz (1 to 2) kHz (2 to 10) kHz	830 μV/V + 11 mV 1000 μV/V + 11 mV 2 % + 11 mV	
(121.2 to 1050) V, Auto, 1 MΩ	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	90 μV/V + 2.6 mV 90 μV/V + 2.6 mV 210 μV/V + 2.6 mV 510 μV/V + 110 mV	

Parameter/Range	Frequency	CMC ^{2, 5} (±)	Comments
AC Voltage ³ – Generate			
(1 to 2.2) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	190 μV/V + 7.0 μV 84 μV/V + 7.0 μV 71 μV/V + 6.0 μV 160 μV/V + 6.0 μV 460 μV/V + 15 μV 850 μV/V + 30 μV 1100 μV/V + 30 μV 3800 μV/V + 30 μV	Fluke 5730A
(2.2 to 22) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	180 μV/V + 4.0 μV 70 μV/V + 4.0 μV 62 μV/V + 4.0 μV 160 μV/V + 4.0 μV 400 μV/V + 5.0 μV 820 μV/V + 10 μV 1100 μV/V + 20 μV 2600 μV/V + 20 μV	
22 to 220 mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	190 μV/V + 12 μV 72 μV/V + 7.0 μV 48 μV/V + 7.0 μV 95 μV/V + 7.0 μV 250 μV/V + 17 μV 510 μV/V + 20 μV 1100 μV/V + 25 μV 2800 μV/V + 45 μV	
220 mV to 2.2 V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	190 μV/V + 40 μV 72 μV/V + 15 μV 36 μV/V + 8 μV 54 μV/V + 10 μV 79 μV/V + 30 μV 260 μV/V + 80 μV 780 μV/V + 200 μV 1400 μV/V + 300 μV	

Parameter/Range	Frequency	CMC ^{2,5} (±)	Comments
AC Voltage ³ – Generate (cont)			
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	190 μV/V + 400 μV 70 μV/V + 150 μV 33 μV/V + 50 μV 52 μV/V + 100 μV 64 μV/V + 200 μV 380 μV/V + 600 μV 840 μV/V + 2000 μV 3900 μV/V + 3200 μV	Fluke 5730A
(22 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	190 μV/V + 4.0 mV 72 μV/V + 1.5 mV 44 μV/V + 0.6 mV 64 μV/V + 1.0 mV 120 μV/V + 2.5 mV 700 μV/V + 16 mV 3400 μV/V + 40 mV 6200 μV/V + 80 mV	
(220 to 1100) V	15 Hz to 50 Hz 50 Hz to 1 kHz	230 μV/V + 16 mV 60 μV/V + 3.5 mV	
(220 to 750) V	(30 to 50) kHz (50 to 100) kHz	470 μV/V + 11 μV 1800 μV/V + 45 μV	Fluke 5725A Amplifier
(750 to 1100) V	(1 to 20) kHz (20 to 30) kHz	130 μV/V + 6 μV 470 μV/V + 11 μV	

Parameter/Range	Frequency	CMC ^{2,5} (±)	Comments
AC Current ³ – Measure			
Up to 20.2 µA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	2100 µA/A + 5 nA 2100 µA/A + 5 nA 2100 µA/A + 5 nA	Fluke 8588A
(20.2 to 202) µA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	260 µA/A + 10 nA 510 µA/A + 0.1 µA 720 µA/A + 1 µA 4100 µA/A + 2 nA	
(0.202 to 2.02) mA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	260 µA/A + 0.1 µA 510 µA/A + 0.1 µA 720 µA/A + 0.1 µA 4100 µA/A + 0.2 µA	
(2.02 to 20.20) mA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	260 µA/A + 1 µA 510 µA/A + 1 µA 720 µA/A + 1 µA 4100 µA/A + 2 µA	
(20.2 to 202) mA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	260 µA/A + 10 µA 500 µA/A + 10 µA 710 µA/A + 10 µA	
(0.202 to 2.02) A	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	260 µA/A + 210 µA 510 µA/A + 210 µA 710 µA/A + 210 µA	
(2.02 to 20.2) A	10 Hz to 2 kHz (2 to 10) kHz	810 µA/A + 1 mA 810 µA/A + 1 mA	
(20.2 to 30.2) A	10 Hz to 2 kHz (2 to 10) kHz	800 µA/A + 130 mA 1200 µA/A + 130 mA	

Parameter/Range	Frequency	CMC ^{2,5} (±)	Comments
AC Current ³ – Generate			
(22 to 220) μA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	210 μA/A + 16 μA 140 μA/A + 10 μA 80 μA/A + 8 μA 220 μA/A + 12 μA 850 μA/A + 65 μA	Fluke 5730A
(.22 to 2.2) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	190 μA/A + 40 μA 120 μA/A + 35 μA 80 μA/A + 35 μA 160 μA/A + 110 μA 850 μA/A + 650 μA	
(2.2 to 22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	190 μA/A + 400 μA 120 μA/A + 350 μA 80 μA/A + 350 μA 160 μA/A + 550 μA 850 μA/A + 5000 μA	
(22 to 220) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	200 μA/A + 4 μA 130 μA/A + 3.5 μA 80 μA/A + 2.5 μA 160 μA/A + 3.5 μA 850 μA/A + 10 μA	
(0.22 to 2.2) A	40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	190 μA/A + 35 μA 350 μA/A + 80 μA 5400 μA/A + 160 μA	
(2.2 to 11) A	40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	360 μA/A + 35 μA 740 μA/A + 80 μA 2800 μA/A + 160 μA	

Parameter/Range	Frequency	CMC ^{2,5} (±)	Comments
AC Current ³ – Generate (cont)			
(12 to 30.2) A	(3 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz	780 μA/A + 10 mA 540 μA/A + 8 mA 3900 μA/A + 8 mA	Fluke 5560A
(30.2 to 60) A	(45 to 65) Hz (65 to 440) Hz	0.64 % output + 0.03 A 1.0 % output + 0.03 A	Fluke 5500A coil & Fluke 5560A
(60 to 155) A	(45 to 65) Hz (65 to 440) Hz	0.66 % output + 0.25 A 1.0 % output + 0.25 A	
(155 to 600) A	(45 to 65) Hz (65 to 440) Hz	0.62 % output + 0.25 A 1.0 % output + 0.25 A	
(600 to 1510) A	(45 to 65) Hz (65 to 440) Hz	0.60 % output + 0.90 A 1.0 % output + 0.90 A	
Capacitance ³ – Generate	Up to 1.2 nF (1.2 to 12) nF (12 to 120) nF 120 nF to 1.2 μF (1.2 to 12) μF (12 to 120) μF 120 μF to 1.2 mF (1.2 to 12) mF (12 to 120) mF	0.23 % + 2 pF 0.24 % + 5 pF 0.09 % + 30 pF 0.08 % + 300 pF 0.09 % + 3 nF 0.09 % + 25 nF 0.15 % + 250 nF 0.15 % + 3 μF 0.30 % + 30 μF	Fluke 5560A

II. Fluid Quantities

Parameter/Equipment	Range	CMC ^{2,6} (±)	Comments
Piston Operated Volumetric Apparatus	(2 to 10) μl (10 to 50) μl (50 to 200) μl (200 to 1000) μl	0.058 μl 0.10 μl 0.26 μl 2.4 μl	Pipettes Gravimetric method

III. Mechanical

Parameter/Equipment	Range	CMC ^{2,4,6} (±)	Comments
Pressure ³ – Measuring Equipment	(Up to 100) psia (>100 to 300) psia (>300 to 1000) psia	0.04 psia 0.14 psia 0.17 psia	Fluke PPC4 pressure module
Scales & Balances ³	>100 mg to 20 g (>20 to 250) g >250 g to 1 kg (>1 to 3) kg (>3 to 10) kg (>10 to 20) kg (>20 to 24) kg	0.03 mg + 0.6R 0.11 mg + 0.6R 0.70 mg + 0.6R 2.0 mg + 0.6R 4.4 mg + 0.6R 16 mg + 0.6R 26 mg + 0.6R	Class 1 weights

IV. Thermodynamics

Parameter/Equipment	Range	CMC ^{2,4,6} (±)	Comments
Humidity ³ – Measuring Equipment	(10 to 20) % RH (>20 to 50) % RH (>50 to 80) % RH	0.13 % RH + 0.6R 0.36 % RH + 0.6R 0.76 % RH + 0.6R	Thunder Scientific 2500
Humidity ³ – Measure	(10 to 35) % RH (35 to 50) % RH (50 to 80) % RH	0.61 % RH 0.63 % RH 1.1 % RH	Rotronics HC2A-3
Temperature ³ – Measure	-197 °C (-197 to 0) °C (0 to 231) °C (231 to 420) °C	3 mK 2 mK 4.5 mK 6.4 mK	SPRT
Temperature – Measuring Equipment ³	(-90 to 140) °C	0.06 °C + 0.6R	SPRT, Fluke Superthermometer readout & dry block
Temperature – Measuring Equipment Triple Point Water Zinc	0.010 °C 419.527 °C	0.002 °C 0.0061 °C	Fluke Mini Fixed Reference Point Cells Fluke Superthermometer Readout

V. Time & Frequency

Parameter/Equipment	Range	CMC ^{2, 6} (\pm)	Comments
Frequency ³ – Measure	(1 to 350) MHz	1.2 μ Hz/Hz	Frequency counter Keysight 53220A
Frequency ³ – Measuring Equipment	0.01 Hz to 2 MHz	3.3 μ Hz/Hz	Fluke 5560A

¹ This laboratory offers commercial and field calibration service.

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ In the statement of Calibration and Measurement Capability, R represents the numerical value of the resolution of the device. In the statement of CMC, percentages are to be read as percent of reading, unless otherwise noted. Percent "FS" represents percent of "Full Scale."

⁵ The stated measure values are determined using the indicated instruments (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

⁶ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

⁷ This scope meets A2LA's *P112 Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

ELLAB, INC. *Doylestown, PA*

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSLI Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 16th day of May 2024.

A blue ink signature of Mr. Trace McInturff, written over a horizontal line.

Mr. Trace McInturff Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 4828.01
Valid to January 31, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.